

Beating & Freeness Tester



Model	FS-150
Standard number	ISO 5267/1 , ISIRI 3789-1
Pulp Volume	1000 ml
Pulp Sample consistency	0.20%
Pulp Temperature adjusted at	20 oC ± 0.05 oC
Lifting method	Penumatic
Reproducibility	within 4%
Weight	40 Kg
Air Pressure	4~6 kg/cm2